T247-EN.01

Human Anti-SARS-CoV-2 (LP.8.1) Antibody IgG Titer Serologic Assay Kit (Spike Trimer)

Pack Size: 96 tests

Catalog Number: RAS-T247

IMPORTANT: Please carefully read this manual before performing your experiment.

For Research Use Only. Not For Use In Diagnostic Or Therapeutic Procedures

INTENDED USE

The kit is developed for qualitative detection and antibody titer determination of Anti-SARS-CoV-2 Spike Trimer (LP.8.1) antibody IgG in human serum. It is intended for research use only (RUO).

PRINCIPLE OF THE ASSAY

The newly identified Severe Acute Respiratory Syndrome Coronavirus 2 (SARS-CoV-2) has posed a serious threat to human health. A rapid and effective Assay kit detecting the levels of Anti-SARS-CoV-2 in human serum can facilitate research on characterization of antibodies produced in response to SARS-CoV-2 infection.

This assay kit is used to measure the titer of Anti-SARS-CoV-2 Spike Trimer (LP.8.1) antibody IgG by employing an indirect ELISA. Immobilize SARS-CoV-2 Spike Trimer (LP.8.1/Omicron) on the microplate. Then add the samples, incubate and wash the wells. Next add Secondary antibody HRP-Conjugated Antibody to the plate, incubate and wash the wells. Lastly load the substrate into the wells and monitor color development in proportion with the amount of antibody present. The reaction is stopped by the addition of a stop solution and the intensity of the absorbance can be measured at 450 nm and 630 nm. The OD Value reflects the amount of antibody bound.

US and Canada: Tel: +1 800-810-0816 | Web: http://www.acrobiosystems.com
Asia and Pacific: Tel: +86 400-682-2521 | E-mail: order@acrobiosystems.com

MATERIALS PROVIDED

TABLE 1. MATERIALS PROVIDED

Catalog	Components	Size (96 tests)	Format	Storage	
				Unopened	Opened
RAS247-C01	Pre-coated SARS-CoV-2 Spike Trimer (LP.8.1/Omicron) Microplate	1 plate	Solid	2-8℃	2-8℃
RAS247-C02	SARS-CoV-2 Antibody Positive Control	100 µL	Liquid	2-8℃	2-8℃
RAS247-C03	SARS-CoV-2 Antibody Negative Control	100 µL	Liquid	2-8℃	2-8℃
RAS247-C04	HRP-Conjugated Antibody	50 µL	Liquid	2-8℃, avoid light	2-8℃, avoid light
RAS247-C05	10×Washing Buffer	50 mL	Liquid	2-8℃	2-8℃
RAS247-C06	Dilution Buffer	50 mL	Liquid	2-8℃	2-8℃
RAS247-C07	Substrate Solution	12 mL	Liquid	2-8℃, avoid light	2-8℃, avoid light
RAS247-C08	Stop Solution	7 mL	Liquid	2-8℃	2-8℃

Note: *It is recommended that HRP-Conjugated Antibody be centrifuged briefly before use to deposit liquid from the tube wall or cap to the bottom of the tube.*

STORAGE AND VALIDITY INSTRUCTIONS

1. Store the unopened kit at 2-8°C upon receipt.

2. Locate the expiration date on the outer packaging and do not use reagents beyond their expiration date.

3. The opened kit should be stored according to the storage conditions listed in the Components Table. The shelf life of the opened kit is 30 days from the date of opening.

MATERIALS REQUIRED BUT NOT PROVIDED

Single or dual wavelength microplate reader with 450 nm and 630 nm filter;

Centrifuge;

37 ºC Incubator;

US and Canada: Tel: +1 800-810-0816
Asia and Pacific: Tel: +86 400-682-2521
Web: http://www.acrobiosystems.com
E-mail: order@acrobiosystems.com

Single channel or multichannel pipettes with 10 µL, 300 µL and 1000 µL precision;

10 µL, 200 µL and 1000 µL pipette tips;

Test Tubes;

Graduated cylinder;

Deionized or distilled water for dilution;

REAGENT PREPARATION

Bring all reagents and samples to room temperature (20℃-25℃) before use.

RECOMMENDED SAMPLE PREPARATION

1. Working fluid preparation

1.1 Preparation of 1×Washing Buffer:

Dilute 50 mL 10×Washing Buffer with ultrapure water/deionized water to 500 mL.

1.2 Preparation of Positive Control and Negative Control working fluid and pre-treatment of samples:

a. For qualitative detection of antibodies:

Dilute the Samples, Positive Control and Negative Control at 1:100 with Dilution Buffer.

b. For determination of antibody titer:

It is recommended to dilute the samples, Positive Control from 1:1600-1:102400 with Dilution Buffer, Negative Control at 1:100 with Dilution Buffer.

2. Plate set up

Number the diluted samples corresponding to the wells of the Pre-coated SARS-CoV-2 Spike Trimer (LP.8.1/Omicron) Microplate. Each experiment requires a set of Positive Control and Negative Control working solution.

3. Add Samples

Add 100 µL diluted Samples, Positive Control and Negative Control working solution to the corresponding wells. Add 100 µL Dilution Buffer to blank control. Seal the plate with microplate sealing film and incubate at 37℃ for 1.0 h.

4. Washing

Carefully remove the plate sealer and discard the liquid from the wells. Add 300 µL of 1×Washing Buffer to each well and soak for 30 seconds. Aspirate or decant the buffer. Repeat for a total of three washes. After the final wash, invert

US and Canada: Tel: +1 800-810-0816 | Web: http://www.acrobiosystems.com
Asia and Pacific: Tel: +86 400-682-2521 | E-mail: order@acrobiosystems.com

the plate and blot dry on absorbent paper.

5. Add HRP-Conjugated Antibody

Dilute **HRP-Conjugated Antibody** solution at 1:1000 with Dilution Buffer to make a working solution. The prepared working solution should be protected from light.

For all wells, add 100 µL HRP-Conjugated Antibody working solution. Seal the plate with microplate sealing film and incubate at 37℃ for 1.0 h, avoid light.

6. Washing

Repeat step 4.

7. Substrate Reaction

Add 100 µL **Substrate Solution** to each well. Seal the plate with microplate sealing film and incubate at 37℃ for 20 minutes, avoid light.

8. Termination

Add 50 µL **Stop Solution** to each well. Gently tap the plate to ensure thorough mixing.

Note: *The color in the wells will change from blue to yellow.*

9. Data Recording

Read the absorbance at 450 nm and 630 nm using UV/Vis microplate spectrophotometer within 5 minutes after adding the stop solution.

Note*: Subtracting the OD630nm value from the OD450nm value helps reduce background interference.*

CUT-OFF VALUE IDENTIFICATION

Cut-off value =0.170

Normal range of Negative control (1:100): $OD_{450\ nm}$-$OD_{630\ nm}$<0.1

Normal range of Positive control (1:1600): $OD_{450\ nm}$-$OD_{630\ nm}$>1.5

Note: *The cut-off value can be determined by the end user.*

INTERPRETION OF RESULTS

a. For qualitative detection of antibodies:

Positive reading: $OD_{450\ nm}$-$OD_{630\ nm}$ of sample ≥ Cut-off value means Anti-SARS-CoV-2 Spike Trimer (LP.8.1) antibody

US and Canada: Tel: +1 800-810-0816 | Web: http://www.acrobiosystems.com
Asia and Pacific: Tel: +86 400-682-2521 | E-mail: order@acrobiosystems.com

IgG are detected.

Negative reading: $OD_{450\ nm}$-$OD_{630\ nm}$ of sample < Cut-off value means Anti-SARS-CoV-2 Spike Trimer (LP.8.1) antibody IgG are not detected.

b. For determination of antibody titer:

Determination of antibody titer: The positive sample is serially diluted. The antibody titer of the sample is defined as the highest dilution factor that still produces a positive signal.

LIMITATIONS OF THE PROCEDURE

The kit cannot be used for quantitative detection.

PRECAUTIONS

1. This kit is for research use only and is not for use in diagnostic or therapeutic applications.
2. This kit should be used according to the provided instructions.
3. Do not mix or interchange reagents from different lots or kits.
4. Bring all reagents and samples to room temperature (20℃-25℃) before use. If crystals have formed in the buffer solution, incubate until the crystals have completely dissolved. Before use, bring the solution back to room temperature.
5. This kit should be stored at 2°C-8°C.
6. Please prepare the working solution of each component according to the needs of the experiment, all prepared working solutions are for one-time use and cannot be stored.

TYPICAL DATA

Note: *The Typical data is for reference only.*

a. For qualitative detection of antibodies:

Value Result in units	Result	Test Result Interpretation
$OD_{450\ nm}$ - $OD_{630\ nm}$ =0.063	Negative	Anti-SARS-CoV-2 Spike Trimer (LP.8.1) antibody IgG are not detected
$OD_{450\ nm}$ - $OD_{630\ nm}$ =0.615	Positive	Anti-SARS-CoV-2 Spike Trimer (LP.8.1) antibody IgG are detected

b. For determination of antibody titer:

Note: *Quality control data between different plates should not be mixed, and negative and positive controls should be set for each test.*

US and Canada: Tel: +1 800-810-0816 | Web: http://www.acrobiosystems.com
Asia and Pacific: Tel: +86 400-682-2521 | E-mail: order@acrobiosystems.com

Ratio of Dilution	$OD_{450\ nm}$ - $OD_{630\ nm}$ (**Samples**)	Result
400	3.188	The titer level of antibody is 51200
800	3.189	
1600	2.691	
3200	1.923	
6400	1.124	
12800	0.678	
25600	0.334	
51200	0.204	
102400	0.125	
204800	0.077	
Blank	0.025	

US and Canada: **Tel**: +1 800-810-0816
Asia and Pacific: **Tel**: +86 400-682-2521
Web: http://www.acrobiosystems.com
E-mail: order@acrobiosystems.com